א.ב. מולי סחר ושרותים בע"מ
A.B. **Moli** Trade & Services Ltd.

Ozone Check

for Water Quality Testing (481234)

SenSafe™

(US Patent # 6541269)

SenSafe Ozone Check for water quality testing is a unique test strip that is designed to test for levels of Ozone in water. Reporting levels at 0, 0.05, 0.1, 0.2, 0.3, 0.4, > 0.5 ppm (mg/L), SenSafe Ozone Check gives you accurate results in only 30 seconds. The low detection ability with this strip is possible because of the unique aperture feature.

The test procedure is as follows: Dip one test strip into a 50mL or larger water sample with constant, gentle back-and-forth motion for 10 seconds. Remove the strip, shake once, briskly, to remove excess water and wait 20 seconds. Then, view through the aperture to match with the closest color. The simple dip-and-read procedure makes testing easy, as no technical experience is required to run the test. Low-range and expanded range versions of the SenSafe Ozone Check are also available.

View the 10 Second Temperature Compensation Chart

Detailed Specifications:

Part #	Number of Tests	Sensitivity	Total Test Time
481234 larger picture	Bottle of 50	0, 0.05, 0.1, 0.2, 0.3, 0.4, > 0.5 ppm (mg/L)	30 Seconds
481134 larger picture	30 Pkts of 1	0, 0.05, 0.1, 0.2, 0.3, 0.4, > 0.5 ppm (mg/L)	30 Seconds

POB 1397, KFAR-SABA 44113, ISRAEL
FAX 972 9 7440681 פקס
HTTP://WWW.ABMOLI.COM

ת.ד. 1397, כפר-סבא 44113
TEL 972 9 7434264 טל.
E-mail assaelm@abmoli.com דוא"ל

א.ב. מולי סחר ושרותים בע"מ
A.B. **Moli** Trade & Services Ltd.

Temperature Compensation Chart

Because the temperature of the tested water will effect the result, you should adjust the dip time according the following chart:

°C / °F	Dip Time
0 / 32	40
1 / 34	38
2 / 36	36
3 / 37	34
4 / 39	32
5 / 41	30
6 / 43	28
7 / 45	27
8 / 46	26
9 / 48	25
10 / 50	24
11 / 52	23
12 / 54	22
13 / 55	20
14 / 57	19
15 / 59	18
16 / 60	17
17 / 62	16
18 / 64	15
19 / 66	14
20 / 68	13
21 / 70	12

°C / ° F	Dip Time
22 / 72	11
23 / 73	10
24 / 75	10
25 / 77	10
26 / 79	9
27 / 81	9
28 / 82	9
29 / 84	8
30 / 86	8
31 / 88	8
32 / 90	8
33 / 91	8
34 / 93	7
35 / 95	7
36 / 97	7
37 / 99	7
38 / 100	7
39 / 102	7
40 / 104	6

Note: Measure the temperature of the water sample within ±2°C / ±4°F

© 2002, ITS, Inc. All Rights Reserved

POB 1397, KFAR-SABA 44113, ISRAEL
FAX 972 9 7440681 פקס
HTTP://WWW.ABMOLI.COM
ת.ד. 1397, כפר-סבא 44113
TEL 972 9 7434264 טל.
E-mail assaelm@abmoli.com דוא"ל